
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000804
11:28 04/10/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:12
PRESSURIZING 3:15
STERILIZING 4:26
136.4 C 233kPa 4:26
136.4 C 233kPa 4:56
136.4 C 233kPa 5:26
136.4 C 231kPa 5:56
136.3 C 233kPa 6:26
136.3 C 232kPa 6:56
136.1 C 230kPa 7:26
136.4 C 232kPa 7:56
136.4 C 232kPa 7:56
VENTING 7:56
AIR DRYING 8:51
CYCLE COMPLETE 8:54

Digital Signature #
D27757E7695AFB31
