
Midmark M11 - v1.0.5
Total Cycles: 1627

Name: RITTER2
Sterilizer ID: V2313228
10/20/2022 06:51 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 78.9 F 0.0
2:00 86.8 F 0.0
4:00 88.4 F 0.0
6:00 106.5 F 0.1
8:00 142.8 F 0.2
10:00 178.7 F 1.5
12:00 201.3 F 4.8
14:00 217.8 F 8.3
16:00 228.9 F 12.1
18:00 228.5 F 7.6
20:00 241.0 F 12.3
22:00 236.7 F 9.7
24:00 237.3 F 9.4
26:00 246.9 F 14.2
28:00 258.1 F 20.3
30:00 268.5 F 26.8

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.0
1:00 272.4 F 29.8
2:00 271.6 F 29.0
3:00 271.9 F 29.6
4:00 272.5 F 29.5

Min 271.1 F 28.8
Max 272.8 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.2 F 0.0
5:00 224.6 F 0.0
10:00 223.1 F 0.0
15:00 227.7 F 0.0
20:00 227.4 F 0.0
25:00 229.2 F 0.0
30:00 224.6 F 0.0
Drying Complete

FILLING: 1:02
HEATING: 31:19
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:09:31
CYCLE COMPLETE

10/20/2022 08:00 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____