

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 002651  
7:46 07/04/2025  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:52  
PRESSURIZING 2:57  
STERILIZING 4:08  
136.0 C 220kPa 4:08  
135.9 C 221kPa 4:38  
136.3 C 224kPa 5:08  
136.4 C 223kPa 5:38  
135.9 C 222kPa 6:08  
136.4 C 223kPa 6:38  
136.2 C 222kPa 7:08  
136.4 C 224kPa 7:38  
136.4 C 224kPa 7:38  
VENTING 7:38  
AIR DRYING 8:31  
CYCLE COMPLETE 9:33  
-----

Digital Signature #  
5AFFC0A29D70DD26  
-----