
Midmark M11 - v1.0.0
Total Cycles: 1502

Name: RITTER 1
Sterilizer ID: V2178198
01/10/2022 09:21 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

mm:ss	Degrees	PSI
0:00	72.6 F	0.0
2:00	82.4 F	0.0
4:00	91.4 F	0.0
6:00	121.1 F	0.0
8:00	147.1 F	0.0
10:00	178.7 F	0.0
12:00	198.8 F	0.0
14:00	210.4 F	0.1
16:00	212.8 F	0.1
18:00	213.0 F	0.1
20:00	213.1 F	0.1
22:00	213.1 F	0.1
24:00	213.3 F	0.2
26:00	215.5 F	0.9
28:00	220.7 F	2.5

C633: HEATUP MODE
PRESSURE LOW
CHAMBER TEMP 222F
CHAMBER PRESS 3.0PSI
** ITEMS NOT STERILE **

01/10/2022 09:51 AM