

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 001509  
10:02 18/10/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:08  
PRESSURIZING 2:14  
STERILIZING 3:10  
136.4 C 223kPa 3:10  
136.0 C 222kPa 3:40  
136.4 C 226kPa 4:10  
136.1 C 222kPa 4:40  
136.4 C 222kPa 5:10  
136.4 C 225kPa 5:40  
136.2 C 224kPa 6:10  
136.0 C 222kPa 6:40  
136.0 C 222kPa 6:40  
VENTING 6:40  
AIR DRYING 7:32  
CYCLE COMPLETE 8:05  
-----

Digital Signature #  
74662350C204EC38  
-----