
Midmark M11 - v1.0.5
Total Cycles: 1017

Name: RITTER2
Sterilizer ID: V2313228
04/21/2022 09:23 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 177.4 F 0.0
2:00 175.9 F 0.7
4:00 196.8 F 4.9
6:00 217.8 F 10.3
8:00 226.5 F 6.7
10:00 240.0 F 12.2
12:00 239.6 F 10.5
14:00 238.1 F 9.7
16:00 252.7 F 16.9
18:00 264.8 F 24.0

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 28.6
1:00 272.3 F 29.4
2:00 272.3 F 28.8
3:00 271.7 F 29.0
4:00 272.5 F 29.2

Min 271.3 F 28.4
Max 272.8 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.5 F 0.0
5:00 226.5 F 0.0
10:00 228.3 F 0.0
15:00 234.8 F 0.0
20:00 204.1 F 0.0
Dry Cycle Aborted

FILLING: 1:40
HEATING: 19:42
STERILIZING: 4:00
VENTING: 3:11
DRYING: 22:50
TOTAL CYCLE: 00:51:24

04/21/2022 10:14 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____