
Midmark M11 - v1.0.0
Total Cycles: 1584

Name: RITTER 1
Sterilizer ID: V2178198
02/01/2022 03:31 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 162.3 F 0.0
2:00 172.5 F 3.1
4:00 179.4 F 7.5
6:00 212.1 F 12.3
8:00 219.4 F 5.7
10:00 231.1 F 10.8
12:00 232.0 F 7.2
14:00 245.0 F 13.4
16:00 243.8 F 12.0
18:00 256.3 F 18.6
20:00 267.7 F 25.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.2
1:00 272.6 F 28.8
2:00 271.5 F 28.1
3:00 272.7 F 29.0
4:00 272.0 F 28.4

Min 271.5 F 28.1
Max 272.8 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.6 F 0.0
5:00 231.4 F 0.0
10:00 243.2 F 0.0
15:00 228.7 F 0.0
20:00 226.2 F 0.0
25:00 230.3 F 0.0
30:00 228.7 F 0.0
Drying Complete

FILLING: 1:18
HEATING: 21:22
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:59:53
CYCLE COMPLETE

02/01/2022 04:30 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____