
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002549
12:35 28/02/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:01
PRESSURIZING 2:08
STERILIZING 3:17
136.0 C 222kPa 3:17
136.0 C 222kPa 3:47
136.4 C 226kPa 4:17
136.4 C 226kPa 4:47
136.4 C 224kPa 5:17
136.1 C 222kPa 5:47
136.4 C 225kPa 6:17
136.2 C 223kPa 6:47
136.2 C 223kPa 6:47
VENTING 6:47
AIR DRYING 7:38
CYCLE COMPLETE 7:57

Digital Signature #
91317477B0422845
