

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1690

Name: RITTER 1  
Sterilizer ID: V2178198  
03/02/2022 09:18 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 159.4 F 0.0  
2:00 160.6 F 1.6  
4:00 166.9 F 4.4  
6:00 189.9 F 8.7  
8:00 215.9 F 9.0  
10:00 222.5 F 8.1  
12:00 235.7 F 13.2  
14:00 236.7 F 10.0  
16:00 236.1 F 8.3  
18:00 248.5 F 14.5  
20:00 260.4 F 21.2  
22:00 271.3 F 28.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 28.4  
1:00 271.7 F 28.7  
2:00 272.4 F 29.0  
3:00 271.7 F 28.4  
4:00 271.6 F 28.6

Min 271.2 F 28.2  
Max 272.6 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 205.9 F 0.0  
5:00 233.7 F 0.0  
10:00 226.8 F 0.0  
15:00 226.0 F 0.0  
20:00 224.4 F 0.0  
25:00 232.4 F 0.0  
30:00 231.9 F 0.0  
Drying Complete

FILLING: 1:02  
HEATING: 22:34  
STERILIZING: 4:00  
VENTING: 3:07  
DRYING: 30:00  
TOTAL CYCLE: 01:00:43  
CYCLE COMPLETE

03/02/2022 10:18 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_