
Midmark M11 - v1.0.5
Total Cycles: 307

Name: RITTER2
Sterilizer ID: V2313228
08/25/2021 06:37 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 72.2 F 0.0
2:00 85.2 F 0.0
4:00 132.5 F 0.0
6:00 157.2 F 0.2
8:00 166.1 F 0.6
10:00 180.6 F 4.4
12:00 195.7 F 8.2
14:00 210.3 F 11.9
16:00 224.0 F 5.6
18:00 231.2 F 9.9
20:00 238.7 F 9.9
22:00 242.1 F 11.8
24:00 239.3 F 10.2
26:00 252.0 F 16.6
28:00 262.6 F 22.8
30:00 272.1 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.7
1:00 272.6 F 29.7
2:00 272.1 F 28.9
3:00 271.4 F 29.1
4:00 272.3 F 29.5

Min 271.1 F 28.5
Max 272.9 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 216.5 F 0.0
5:00 218.6 F 0.0
10:00 206.1 F 0.0
15:00 194.5 F 0.0
20:00 208.8 F 0.0
25:00 207.9 F 0.0
30:00 211.5 F 0.0
Drying Complete

FILLING: 1:00
HEATING: 30:29
STERILIZING: 4:00
VENTING: 3:00
DRYING: 30:00
TOTAL CYCLE: 01:08:30
CYCLE COMPLETE

08/25/2021 07:59 AM
ACCEPTED: X
REJECTED:
Operator: U02