
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 008140
15:01 03/12/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:00
PRESSURIZING 3:02
STERILIZING 4:06
136.0 C 220kPa 4:06
136.4 C 222kPa 4:36
136.2 C 223kPa 5:06
136.2 C 221kPa 5:36
136.4 C 220kPa 6:06
136.5 C 223kPa 6:36
136.4 C 223kPa 7:06
136.5 C 222kPa 7:36
136.5 C 222kPa 7:36
VENTING 7:36
AIR DRYING 8:32
CYCLE COMPLETE 8:41

Digital Signature #
A7CDE4EC4E37D4AD
