
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 000052
10:28 06/03/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:56
PRESSURIZING 3:03
STERILIZING 4:02
136.0 C 220kPa 4:02
136.4 C 226kPa 4:32
136.4 C 224kPa 5:02
136.1 C 223kPa 5:32
136.4 C 222kPa 6:02
136.1 C 223kPa 6:32
136.2 C 224kPa 7:02
136.4 C 225kPa 7:32
136.4 C 225kPa 7:32
VENTING 7:32
AIR DRYING 8:26
CYCLE COMPLETE 23:26

Digital Signature #
0E76E02973796745
