

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1203

Name: RITTER2  
Sterilizer ID: V2313228  
06/21/2022 06:47 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 74.7 F 0.0  
2:00 75.0 F 0.0  
4:00 77.8 F 0.0  
6:00 102.0 F 0.0  
8:00 141.0 F 1.5  
10:00 167.1 F 4.7  
12:00 181.6 F 8.3  
14:00 200.7 F 12.6  
16:00 219.2 F 6.1  
18:00 230.1 F 10.3  
20:00 237.0 F 9.0  
22:00 240.7 F 11.9  
24:00 239.0 F 10.2  
26:00 251.5 F 16.5  
28:00 262.5 F 23.0  
30:00 272.8 F 29.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.9  
1:00 272.2 F 29.9  
2:00 271.6 F 29.6  
3:00 271.4 F 29.2  
4:00 272.2 F 29.5

Min 270.6 F 28.7  
Max 273.1 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 174.7 F 0.0  
5:00 219.4 F 0.0  
10:00 197.4 F 0.0  
15:00 190.5 F 0.0  
20:00 171.4 F 0.0  
25:00 173.0 F 0.0  
30:00 167.6 F 0.0  
Drying Complete

FILLING: 1:00  
HEATING: 30:21  
STERILIZING: 4:00  
VENTING: 3:07  
DRYING: 30:00  
TOTAL CYCLE: 01:08:28  
CYCLE COMPLETE

06/21/2022 08:12 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02