
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 004625
14:01 26/10/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:48
PRESSURIZING 1:55
STERILIZING 2:53
136.5 C 223kPa 2:53
135.9 C 221kPa 3:23
136.4 C 223kPa 3:53
136.4 C 225kPa 4:23
136.3 C 222kPa 4:53
136.3 C 223kPa 5:23
136.4 C 221kPa 5:53
136.3 C 223kPa 6:23
136.3 C 223kPa 6:23
VENTING 6:23
AIR DRYING 7:11
CYCLE COMPLETE 20:00

Digital Signature #
F1738A5DA98B5B42
