
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000924
17:07 10/02/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:54
PRESSURIZING 4:01
STERILIZING 5:35
136.4 C 232kPa 5:35
136.4 C 233kPa 6:05
136.2 C 232kPa 6:35
136.3 C 230kPa 7:05
136.4 C 232kPa 7:35
136.3 C 232kPa 8:05
136.3 C 232kPa 8:35
136.4 C 233kPa 9:05
136.4 C 233kPa 9:05
VENTING 9:05
AIR DRYING 10:03
CYCLE COMPLETE 25:03

Digital Signature #
943DA076F7200902
