
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.8uS / 0.5ppm
CYCLE NUMBER 002707
15:32 17/11/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:27
PRESSURIZING 2:33
STERILIZING 3:29
136.4 C 223kPa 3:29
136.5 C 221kPa 3:59
136.4 C 225kPa 4:29
136.2 C 222kPa 4:59
135.9 C 221kPa 5:29
136.4 C 224kPa 5:59
136.4 C 223kPa 6:29
136.2 C 223kPa 6:59
136.2 C 223kPa 6:59
VENTING 6:59
CYCLE ABORTED 7:22
STOP BUTTON PRESSED
NOT STERILE
END TIME 7:22

Digital Signature #
0DD2C2CAA4F2A836
