
Midmark M11 - v1.0.5
Total Cycles: 374

Name: RITTER2
Sterilizer ID: V2313228
09/16/2021 01:19 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 177.1 F 0.0
2:00 174.0 F 3.3
4:00 184.6 F 8.1
6:00 208.0 F 11.3
8:00 228.8 F 9.5
10:00 236.0 F 9.0
12:00 243.1 F 13.2
14:00 242.8 F 12.4
16:00 256.1 F 19.3
18:00 267.3 F 26.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.2
1:00 272.6 F 29.7
2:00 272.0 F 29.6
3:00 272.2 F 29.6
4:00 271.8 F 29.5

Min 271.5 F 29.1
Max 273.4 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.5 F 0.0
5:00 236.7 F 0.0
10:00 203.3 F 0.0
15:00 221.1 F 0.0
20:00 223.4 F 0.0
25:00 202.1 F 0.0
30:00 188.8 F 0.0
Drying Complete

FILLING: 1:46
HEATING: 19:19
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:58:12
CYCLE COMPLETE

09/16/2021 02:17 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____