

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 000654  
10:26 13/09/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:07  
PRESSURIZING 3:11  
STERILIZING 4:18  
136.0 C 229kPa 4:18  
136.3 C 232kPa 4:48  
136.4 C 233kPa 5:18  
136.4 C 232kPa 5:48  
136.4 C 233kPa 6:18  
136.2 C 231kPa 6:48  
136.1 C 231kPa 7:18  
136.4 C 233kPa 7:48  
136.4 C 233kPa 7:48  
VENTING 7:48  
AIR DRYING 8:41  
CYCLE COMPLETE 8:42

-----  
Digital Signature #  
365DF7515F8BE9E3  
-----