
Midmark M11 - v1.0.5
Total Cycles: 2550

Name: RITTER2
Sterilizer ID: V2313228
06/22/2023 08:37 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 161.4 F 0.0
2:00 161.2 F 1.5
4:00 161.0 F 4.7
6:00 189.7 F 9.2
8:00 210.5 F 8.0
10:00 222.6 F 8.3
12:00 235.0 F 13.3
14:00 236.8 F 10.4
16:00 234.9 F 8.2
18:00 248.9 F 15.1
20:00 260.5 F 21.7
22:00 271.1 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.0
1:00 272.5 F 29.9
2:00 272.4 F 29.5
3:00 271.6 F 28.9
4:00 271.4 F 29.1

Min 271.2 F 28.8
Max 272.8 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.8 F 0.0
5:00 209.3 F 0.0
10:00 224.8 F 0.0
15:00 223.9 F 0.0
20:00 228.3 F 0.0
25:00 230.0 F 0.0
30:00 227.9 F 0.0
Drying Complete

FILLING: 1:07
HEATING: 22:49
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:01:08
CYCLE COMPLETE

06/22/2023 09:38 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____