
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
1.0uS / 0.6ppm
CYCLE NUMBER 001593
12:00 11/18/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:08
PRESSURIZING 6:10
STERILIZING 7:52
136.0 C 220kPa 7:52
136.4 C 217kPa 8:22
136.5 C 217kPa 8:52
136.3 C 221kPa 9:23
136.5 C 221kPa 9:53
136.4 C 219kPa 10:23
136.1 C 220kPa 10:53
136.0 C 219kPa 11:23
VENTING 11:23
AIR DRYING 12:31
CYCLE COMPLETE 37:26

Digital Signature #
44A6F0598D22085E
