

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2143

Name: RITTER2  
Sterilizer ID: V2313228  
03/13/2023 02:29 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 184.9 F 0.0  
2:00 181.5 F 3.8  
4:00 198.0 F 8.6  
6:00 217.1 F 9.1  
8:00 232.9 F 10.0  
10:00 232.3 F 8.3  
12:00 241.4 F 11.5  
14:00 246.2 F 13.8  
16:00 259.3 F 20.9  
18:00 270.8 F 28.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 29.5  
1:00 272.8 F 29.8  
2:00 272.6 F 30.2  
3:00 272.2 F 29.5  
4:00 272.6 F 30.2  
  
Min 271.7 F 29.4  
Max 273.1 F 30.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 202.0 F 0.0  
5:00 222.4 F 0.0  
10:00 224.6 F 0.0  
15:00 212.4 F 0.0  
20:00 215.1 F 0.0  
25:00 233.7 F 0.0  
30:00 227.5 F 0.0  
  
Drying Complete

FILLING: 1:58  
HEATING: 19:29  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 30:00  
TOTAL CYCLE: 00:58:42  
CYCLE COMPLETE

03/13/2023 03:27 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_