

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2401

Name: RITTER2  
Sterilizer ID: V2313228  
05/23/2023 01:41 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 144.7 F 0.0  
2:00 144.7 F 2.1  
4:00 155.6 F 5.4  
6:00 184.6 F 9.9  
8:00 214.0 F 5.9  
10:00 224.2 F 9.1  
12:00 233.3 F 10.6  
14:00 238.5 F 11.4  
16:00 237.5 F 9.7  
18:00 250.7 F 16.2  
20:00 262.2 F 22.8  
22:00 272.7 F 29.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 29.0  
1:00 272.4 F 29.8  
2:00 271.5 F 29.0  
3:00 271.5 F 29.4  
4:00 273.0 F 30.2

Min 271.0 F 28.8  
Max 273.0 F 30.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 185.5 F 0.0  
5:00 212.7 F 0.0  
10:00 230.2 F 0.0  
15:00 224.0 F 0.0  
20:00 223.1 F 0.0  
25:00 224.4 F 0.0  
30:00 222.1 F 0.0  
Drying Complete

FILLING: 1:22  
HEATING: 22:24  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 01:00:57  
CYCLE COMPLETE

05/23/2023 02:41 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_