
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002266
9:10 15/11/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:35
CYCLE ABORTED 3:45
CYCLE FAULT 3
NOT STERILE
END TIME 3:45

Digital Signature #
CA611F4F7640559D
