
Midmark M11 - v1.0.0
Total Cycles: 2708

Name: RITTER 1
Sterilizer ID: V2178198
12/21/2022 11:17 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 186.3 F 0.0
2:00 187.2 F 2.8
4:00 205.5 F 7.2
6:00 220.5 F 12.7
8:00 227.1 F 8.1
10:00 239.0 F 12.2
12:00 239.7 F 11.0
14:00 238.3 F 9.4
16:00 252.3 F 16.3
18:00 264.0 F 23.3
20:00 271.6 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.4
1:00 272.7 F 29.0
2:00 271.7 F 28.3
3:00 272.8 F 29.3
4:00 272.0 F 28.6

Min 271.6 F 28.3
Max 272.9 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 214.0 F 0.0
5:00 221.2 F 0.0
10:00 247.3 F 0.0
15:00 247.1 F 0.0
20:00 236.4 F 0.0
25:00 223.0 F 0.0
30:00 222.4 F 0.0
Drying Complete

FILLING: 1:35
HEATING: 20:02
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:58:48
CYCLE COMPLETE

12/21/2022 12:15 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____