
Midmark M11 - v1.0.0
Total Cycles: 4632

Name: RITTER 1
Sterilizer ID: V2178198
06/12/2024 01:59 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 178.6 F 0.0
2:00 173.2 F 3.2
4:00 181.4 F 6.9
6:00 206.1 F 11.5
8:00 222.2 F 5.5
10:00 231.6 F 10.2
12:00 234.0 F 8.1
14:00 242.4 F 12.1
16:00 239.9 F 10.3
18:00 252.0 F 16.3
20:00 262.5 F 22.5
22:00 272.6 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.8
1:00 271.6 F 28.7
2:00 271.8 F 28.8
3:00 272.4 F 29.0
4:00 272.7 F 29.2

Min 271.4 F 28.7
Max 272.8 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.1 F 0.0
5:00 192.9 F 0.0
10:00 251.5 F 0.0
15:00 240.9 F 0.0
20:00 219.6 F 0.0
25:00 216.9 F 0.0
30:00 220.1 F 0.0
Drying Complete

FILLING: 1:22
HEATING: 23:02
STERILIZING: 4:00
VENTING: 3:48
DRYING: 30:00
TOTAL CYCLE: 01:02:12
CYCLE COMPLETE

06/12/2024 03:01 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____