
Midmark M11 - v1.0.5
Total Cycles: 2505

Name: RITTER2
Sterilizer ID: V2313228
06/13/2023 08:35 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 162.9 F 0.0
2:00 165.4 F 2.0
4:00 172.9 F 5.3
6:00 215.5 F 10.0
8:00 218.3 F 5.8
10:00 229.2 F 9.9
12:00 234.6 F 8.3
14:00 240.9 F 12.8
16:00 240.6 F 11.5
18:00 253.7 F 18.1
20:00 265.3 F 25.1

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.4
1:00 272.9 F 30.0
2:00 271.1 F 29.3
3:00 272.6 F 30.2
4:00 272.8 F 30.3

Min 271.0 F 29.3
Max 273.3 F 30.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.2 F 0.0
5:00 215.2 F 0.0
10:00 234.7 F 0.0
15:00 229.8 F 0.0
20:00 231.1 F 0.0
25:00 228.3 F 0.0
30:00 226.8 F 0.0
Drying Complete

FILLING: 1:05
HEATING: 21:47
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:00:05
CYCLE COMPLETE

06/13/2023 09:35 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____