
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000659
13:39 13/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:33
PRESSURIZING 2:42
STERILIZING 3:51
136.0 C 228kPa 3:51
136.3 C 231kPa 4:21
136.3 C 232kPa 4:51
136.1 C 231kPa 5:21
136.3 C 232kPa 5:51
136.2 C 232kPa 6:21
136.4 C 232kPa 6:51
136.3 C 231kPa 7:21
136.3 C 231kPa 7:21
VENTING 7:21
AIR DRYING 8:16
CYCLE COMPLETE 8:17

Digital Signature #
F2A208BF636BD3AA
