
Midmark M11 - v1.0.0
Total Cycles: 3388

Name: RITTER 1
Sterilizer ID: V2178198
06/14/2023 11:47 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 204.5 F 0.0
2:00 197.8 F 1.9
4:00 214.9 F 7.7
6:00 231.5 F 13.5
8:00 234.2 F 9.9
10:00 235.4 F 8.4
12:00 241.7 F 10.9
14:00 247.8 F 14.1
16:00 260.7 F 21.3
18:00 272.4 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.4
1:00 272.1 F 28.6
2:00 272.6 F 29.1
3:00 271.7 F 28.3
4:00 272.6 F 29.0

Min 271.6 F 28.2
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.0 F 0.0
5:00 220.0 F 0.0
10:00 231.7 F 0.0
15:00 233.3 F 0.0
20:00 226.7 F 0.0
25:00 222.0 F 0.0
30:00 218.8 F 0.0
Drying Complete

FILLING: 2:02
HEATING: 18:24
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 00:57:44
CYCLE COMPLETE

06/14/2023 12:44 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____