
Midmark M11 - v1.0.0
Total Cycles: 4412

Name: RITTER 1
Sterilizer ID: V2178198
03/27/2024 01:48 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 142.6 F 0.0
2:00 141.9 F 1.9
4:00 162.3 F 4.9
6:00 197.2 F 8.6
8:00 226.7 F 12.9
10:00 222.7 F 5.6
12:00 232.9 F 9.7
14:00 238.8 F 10.4
16:00 240.1 F 10.8
18:00 235.6 F 8.4
20:00 248.9 F 14.5
22:00 259.7 F 20.5
24:00 269.6 F 26.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.6
1:00 272.1 F 28.6
2:00 272.7 F 28.9
3:00 272.2 F 29.1
4:00 271.3 F 28.6

Min 271.3 F 28.5
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.1 F 0.0
5:00 195.2 F 0.0
10:00 203.4 F 0.0
15:00 202.8 F 0.0
20:00 198.8 F 0.0
25:00 199.2 F 0.0
30:00 192.6 F 0.0
Drying Complete

FILLING: 1:08
HEATING: 25:43
STERILIZING: 4:00
VENTING: 3:49
DRYING: 30:00
TOTAL CYCLE: 01:04:40
CYCLE COMPLETE

03/27/2024 02:52 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____