
Midmark M11 - v1.0.5
Total Cycles: 2349

Name: RITTER2
Sterilizer ID: V2313228
05/10/2023 12:30 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 191.6 F 0.0
2:00 188.3 F 3.5
4:00 190.9 F 8.2
6:00 215.3 F 10.1
8:00 227.3 F 9.7
10:00 232.7 F 8.8
12:00 242.6 F 13.2
14:00 242.5 F 12.2
16:00 255.5 F 18.9
18:00 266.8 F 25.9

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.1
1:00 272.6 F 29.9
2:00 272.0 F 29.3
3:00 271.4 F 29.3
4:00 272.7 F 29.9

Min 271.2 F 28.9
Max 272.7 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.7 F 0.0
5:00 221.0 F 0.0
10:00 227.8 F 0.0
15:00 232.2 F 0.0
20:00 224.0 F 0.0
25:00 223.9 F 0.0
30:00 221.8 F 0.0
Drying Complete

FILLING: 1:55
HEATING: 19:32
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:58:43
CYCLE COMPLETE

05/10/2023 01:28 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____