

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.5uS / 0.3ppm  
CYCLE NUMBER 001654  
16:21 06/12/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:34  
PRESSURIZING 2:43  
STERILIZING 3:49  
136.4 C 224kPa 3:49  
136.3 C 225kPa 4:19  
136.1 C 223kPa 4:49  
136.0 C 220kPa 5:19  
136.1 C 223kPa 5:49  
136.3 C 224kPa 6:19  
136.4 C 226kPa 6:49  
136.4 C 225kPa 7:19  
136.4 C 225kPa 7:19  
VENTING 7:20  
AIR DRYING 8:11  
CYCLE COMPLETE 8:18  
-----

Digital Signature #  
0D67C048678993A8  
-----