
Midmark M11 - v1.0.0
Total Cycles: 5039

Name: RITTER 1
Sterilizer ID: V2178198
10/30/2024 03:11 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 96.5 F 0.0
2:00 98.7 F 0.0
4:00 125.2 F 0.3
6:00 158.6 F 1.0
8:00 186.1 F 3.9
10:00 202.7 F 7.6
12:00 217.0 F 11.4
14:00 220.2 F 4.5
16:00 233.2 F 9.1
18:00 242.4 F 12.7
20:00 240.1 F 10.6
22:00 235.2 F 7.9
24:00 249.2 F 14.6
26:00 260.3 F 20.8
28:00 270.7 F 27.5

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.1
1:00 271.9 F 28.5
2:00 272.3 F 28.5
3:00 271.5 F 28.0
4:00 272.4 F 28.9

Min 271.3 F 27.9
Max 272.6 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.1 F 0.0
5:00 213.8 F 0.0
10:00 190.6 F 0.0
15:00 189.8 F 0.0
20:00 187.7 F 0.0
25:00 188.9 F 0.0
30:00 185.8 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 28:41
STERILIZING: 4:00
VENTING: 3:34
DRYING: 30:00
TOTAL CYCLE: 01:07:04
CYCLE COMPLETE

10/30/2024 04:18 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____