

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1080

Name: RITTER 1  
Sterilizer ID: V2178198  
08/09/2021 01:14 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 132.7 F 0.0  
2:00 142.5 F 0.8  
4:00 165.4 F 3.2  
6:00 180.7 F 6.5  
8:00 200.8 F 10.7  
10:00 213.2 F 5.2  
12:00 223.9 F 8.0  
14:00 234.5 F 12.4  
16:00 234.6 F 8.9  
18:00 241.0 F 10.9  
20:00 244.9 F 12.8  
22:00 256.4 F 18.8  
24:00 266.9 F 25.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.4  
1:00 272.6 F 29.1  
2:00 271.9 F 28.5  
3:00 271.6 F 28.6  
4:00 272.5 F 29.1

Min 271.4 F 28.2  
Max 272.6 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 201.1 F 0.0  
5:00 215.2 F 0.0  
10:00 230.7 F 0.0  
15:00 220.3 F 0.0  
20:00 216.6 F 0.0  
25:00 211.2 F 0.0  
30:00 210.1 F 0.0  
Drying Complete

FILLING: 1:08  
HEATING: 25:36  
STERILIZING: 4:00  
VENTING: 3:21  
DRYING: 30:00  
TOTAL CYCLE: 01:04:05  
CYCLE COMPLETE

08/09/2021 02:20 PM  
ACCEPTED: X  
REJECTED:  
Operator: U01