
Midmark M11 - v1.0.0
Total Cycles: 3432

Name: RITTER 1
Sterilizer ID: V2178198
06/23/2023 07:46 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.8 F 0.0
2:00 90.1 F 0.0
4:00 90.5 F 0.0
6:00 119.4 F 0.1
8:00 153.3 F 1.6
10:00 177.3 F 4.3
12:00 194.0 F 7.4
14:00 210.7 F 10.9
16:00 218.4 F 5.9
18:00 227.3 F 7.8
20:00 236.9 F 11.9
22:00 233.4 F 8.1
24:00 245.5 F 12.8
26:00 242.9 F 11.8
28:00 254.7 F 17.7
30:00 265.1 F 23.8

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.2
1:00 272.1 F 28.8
2:00 271.9 F 28.3
3:00 271.7 F 28.7
4:00 272.2 F 28.6

Min 271.1 F 28.0
Max 272.8 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.2 F 0.0
5:00 216.7 F 0.0
10:00 215.1 F 0.0
15:00 229.3 F 0.0
20:00 222.9 F 0.0
25:00 217.3 F 0.0
30:00 214.5 F 0.0
Drying Complete

FILLING: 0:46
HEATING: 31:50
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:09:49
CYCLE COMPLETE

06/23/2023 08:55 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____