
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
2.5uS / 1.6ppm
CYCLE NUMBER 001866
7:31 18/07/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:42
PRESSURIZING 2:49
STERILIZING 3:49
136.0 C 220kPa 3:49
136.4 C 224kPa 4:19
136.2 C 221kPa 4:49
136.4 C 224kPa 5:19
136.4 C 223kPa 5:49
136.4 C 224kPa 6:19
136.4 C 224kPa 6:49
136.4 C 223kPa 7:19
136.4 C 223kPa 7:19
VENTING 7:19
AIR DRYING 8:10
CYCLE COMPLETE 8:25

Digital Signature #
C87A40F8859BCEDD
