
Midmark M11 - v1.0.5
Total Cycles: 1680

Name: RITTER2
Sterilizer ID: V2313228
11/03/2022 12:46 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 146.5 F 0.0
2:00 147.8 F 2.4
4:00 161.0 F 5.9
6:00 187.2 F 10.7
8:00 213.7 F 4.5
10:00 225.3 F 9.6
12:00 232.2 F 9.2
14:00 240.7 F 12.1
16:00 239.6 F 10.7
18:00 252.7 F 17.2
20:00 264.1 F 24.0
22:00 271.2 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.0
1:00 272.5 F 29.8
2:00 272.0 F 29.1
3:00 271.4 F 29.2
4:00 272.5 F 29.8

Min 271.2 F 28.8
Max 272.6 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.0 F 0.0
5:00 227.0 F 0.0
10:00 238.1 F 0.0
15:00 225.5 F 0.0
20:00 224.2 F 0.0
25:00 223.6 F 0.0
30:00 224.6 F 0.0
Drying Complete

FILLING: 1:14
HEATING: 22:03
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:00:29
CYCLE COMPLETE

11/03/2022 01:46 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____