
Midmark M11 - v1.0.5
Total Cycles: 4328

Name: RITTER2
Sterilizer ID: V2313228
02/03/2025 01:35 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 129.6 F 0.0
2:00 131.4 F 1.4
4:00 140.3 F 4.1
6:00 167.1 F 7.5
8:00 185.6 F 11.5
10:00 210.0 F 4.2
12:00 222.2 F 8.3
14:00 234.5 F 12.6
16:00 233.3 F 9.1
18:00 241.2 F 11.3
20:00 244.2 F 12.9
22:00 255.1 F 18.7
24:00 265.2 F 24.8
26:00 271.2 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.2
1:00 272.1 F 30.1
2:00 273.3 F 30.5
3:00 272.2 F 29.9
4:00 271.8 F 30.1

Min 271.0 F 29.1
Max 273.3 F 30.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.1 F 0.0
5:00 184.7 F 0.0
10:00 206.5 F 0.0
15:00 204.3 F 0.0
20:00 205.3 F 0.0
25:00 202.5 F 0.0
30:00 203.7 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 26:03
STERILIZING: 4:00
VENTING: 3:56
DRYING: 30:00
TOTAL CYCLE: 01:04:56
CYCLE COMPLETE

02/03/2025 02:39 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____