
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 002506
14:45 22/10/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:07
PRESSURIZING 3:08
STERILIZING 4:16
136.5 C 220kPa 4:16
136.2 C 222kPa 4:47
136.0 C 221kPa 5:17
136.4 C 224kPa 5:47
136.2 C 223kPa 6:17
136.3 C 222kPa 6:47
136.4 C 222kPa 7:17
136.3 C 221kPa 7:47
136.3 C 221kPa 7:47
VENTING 7:47
AIR DRYING 8:41
CYCLE COMPLETE 9:19

Digital Signature #
31B3F01D06169DB5
