
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002498
10:39 10/02/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:51
PRESSURIZING 2:52
STERILIZING 3:46
136.0 C 221kPa 3:46
136.4 C 226kPa 4:16
136.4 C 226kPa 4:46
136.1 C 222kPa 5:16
136.3 C 224kPa 5:46
136.4 C 227kPa 6:16
136.2 C 224kPa 6:46
136.3 C 225kPa 7:16
136.3 C 225kPa 7:16
VENTING 7:16
AIR DRYING 8:09
CYCLE COMPLETE 8:35

Digital Signature #
9BD18F5311E90C81
