
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 006688
9:27 05/12/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:20
PRESSURIZING 3:22
STERILIZING 4:22
136.0 C 220kPa 4:22
136.0 C 221kPa 4:52
136.4 C 222kPa 5:22
136.4 C 220kPa 5:52
136.4 C 222kPa 6:22
136.2 C 220kPa 6:52
136.1 C 222kPa 7:22
136.4 C 223kPa 7:52
136.4 C 223kPa 7:52
VENTING 7:52
AIR DRYING 8:43
CYCLE COMPLETE 8:47

Digital Signature #
24509550005FBD84
