

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3139

Name: RITTER 1  
Sterilizer ID: V2178198  
04/20/2023 02:05 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 189.5 F 0.0  
2:00 184.5 F 0.0  
4:00 201.7 F 3.8  
6:00 218.3 F 8.7  
8:00 232.5 F 10.5  
10:00 237.5 F 10.9  
12:00 236.0 F 8.8  
14:00 238.9 F 9.5  
16:00 250.5 F 15.2  
18:00 263.3 F 22.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 27.9  
1:00 272.5 F 28.5  
2:00 271.4 F 28.0  
3:00 272.5 F 28.6  
4:00 271.6 F 27.9  
  
Min 271.4 F 27.8  
Max 272.7 F 28.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 199.8 F 0.0  
5:00 224.2 F 0.0  
10:00 235.7 F 0.0  
15:00 235.1 F 0.0  
20:00 228.1 F 0.0  
25:00 227.6 F 0.0  
30:00 225.9 F 0.0  
Drying Complete

FILLING: 0:55  
HEATING: 19:60  
STERILIZING: 4:00  
VENTING: 3:18  
DRYING: 30:00  
TOTAL CYCLE: 00:58:13  
CYCLE COMPLETE

04/20/2023 03:03 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_