
Midmark M11 - v1.0.0
Total Cycles: 4079

Name: RITTER 1
Sterilizer ID: V2178198
12/04/2023 02:17 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 104.3 F 0.0
2:00 104.2 F 0.9
4:00 118.4 F 3.2
6:00 149.3 F 6.1
8:00 187.6 F 9.5
10:00 205.8 F 9.0
12:00 212.6 F 5.1
14:00 221.2 F 9.0
16:00 232.5 F 13.3
18:00 235.1 F 9.6
20:00 239.3 F 10.2
22:00 245.1 F 13.0
24:00 256.3 F 18.8
26:00 266.6 F 25.2

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.4
1:00 272.3 F 29.3
2:00 272.1 F 28.9
3:00 271.3 F 28.3
4:00 271.6 F 28.8

Min 271.1 F 28.3
Max 272.5 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.5 F 0.0
5:00 233.6 F 0.0
10:00 221.3 F 0.0
15:00 217.6 F 0.0
20:00 217.6 F 0.0
25:00 220.5 F 0.0
30:00 221.6 F 0.0

Drying Complete

FILLING: 0:57
HEATING: 27:41
STERILIZING: 4:00
VENTING: 3:45
DRYING: 30:00
TOTAL CYCLE: 01:06:24
CYCLE COMPLETE

12/04/2023 03:23 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____