
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002633
10:34 28/03/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:02
PRESSURIZING 2:07
STERILIZING 3:06
136.0 C 220kPa 3:06
136.3 C 223kPa 3:36
136.3 C 224kPa 4:06
136.4 C 224kPa 4:36
136.4 C 226kPa 5:06
136.3 C 224kPa 5:36
136.4 C 224kPa 6:06
136.4 C 224kPa 6:36
136.4 C 224kPa 6:36
VENTING 6:36
AIR DRYING 7:28
CYCLE COMPLETE 29:57

Digital Signature #
E3F46D94522B3B57
