
Midmark M11 - v1.0.5
Total Cycles: 2313

Name: RITTER2
Sterilizer ID: V2313228
04/28/2023 10:50 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 195.5 F 0.0
2:00 189.5 F 3.7
4:00 211.8 F 10.1
6:00 223.7 F 6.1
8:00 238.1 F 12.0
10:00 238.9 F 10.5
12:00 238.6 F 9.9
14:00 253.4 F 17.2
16:00 265.7 F 24.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.5
1:00 272.4 F 28.7
2:00 272.6 F 29.3
3:00 271.7 F 28.4
4:00 272.7 F 29.2

Min 271.5 F 28.3
Max 272.8 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.5 F 0.0
5:00 212.9 F 0.0
10:00 226.2 F 0.0
15:00 229.7 F 0.0
20:00 228.6 F 0.0
25:00 228.8 F 0.0
30:00 226.7 F 0.0
Drying Complete

FILLING: 2:33
HEATING: 17:37
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:57:20
CYCLE COMPLETE

04/28/2023 11:47 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____