
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 004095
14:32 25/07/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:27
PRESSURIZING 2:31
STERILIZING 3:30
136.4 C 220kPa 3:30
136.3 C 225kPa 4:00
136.4 C 219kPa 4:30
135.9 C 220kPa 5:00
136.4 C 223kPa 5:30
136.4 C 223kPa 6:00
136.4 C 226kPa 6:30
136.4 C 218kPa 7:00
136.4 C 218kPa 7:00
VENTING 7:00
AIR DRYING 8:56
CYCLE COMPLETE 9:07

Digital Signature #
5A1731F839D51F45
