
Midmark M11 - v1.0.0
Total Cycles: 2634

Name: RITTER 1
Sterilizer ID: V2178198
12/02/2022 11:34 AM

BEGIN UNWRAPPED CYCLE
Temp: 270 Degrees F
Time: 3 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 161.9 F 0.0
2:00 161.5 F 2.5
4:00 166.9 F 6.1
6:00 192.4 F 11.2
8:00 214.7 F 4.6
10:00 231.1 F 10.0
12:00 234.4 F 7.4
14:00 244.8 F 13.0
16:00 242.8 F 11.3
18:00 256.5 F 18.4
20:00 268.5 F 25.8

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 27.9
1:00 271.9 F 27.9
2:00 272.5 F 28.3
3:00 272.9 F 28.8

Min 271.5 F 27.8
Max 273.0 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 228.0 F 0.0
5:00 229.3 F 0.0
10:00 224.8 F 0.0
Dry Cycle Aborted

FILLING: 1:07
HEATING: 21:05
STERILIZING: 3:00
VENTING: 2:50
DRYING: 12:57
TOTAL CYCLE: 00:40:59

12/02/2022 12:14 PM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____