

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3255

Name: RITTER2  
Sterilizer ID: V2313228  
12/27/2023 12:20 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 168.7 F 0.0  
2:00 160.0 F 2.3  
4:00 175.5 F 6.3  
6:00 199.4 F 11.5  
8:00 224.4 F 6.4  
10:00 236.7 F 11.8  
12:00 235.0 F 9.2  
14:00 237.9 F 9.5  
16:00 248.6 F 14.8  
18:00 261.5 F 22.1  
20:00 272.8 F 29.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 28.8  
1:00 272.3 F 29.6  
2:00 271.9 F 28.9  
3:00 271.7 F 29.4  
4:00 272.7 F 29.4

Min 270.9 F 28.6  
Max 273.3 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 190.6 F 0.0  
5:00 230.7 F 0.0  
10:00 220.7 F 0.0  
15:00 197.8 F 0.0  
20:00 181.8 F 0.0  
25:00 179.8 F 0.0  
30:00 184.9 F 0.0  
Drying Complete

FILLING: 1:08  
HEATING: 20:15  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 00:58:36  
CYCLE COMPLETE

12/27/2023 01:18 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_