

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2169

Name: RITTER2  
Sterilizer ID: V2313228  
03/22/2023 01:38 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 167.6 F 0.0  
2:00 167.6 F 3.2  
4:00 183.1 F 7.6  
6:00 207.1 F 13.1  
8:00 223.4 F 8.5  
10:00 234.2 F 10.8  
12:00 239.5 F 12.0  
14:00 239.7 F 10.9  
16:00 253.4 F 17.8  
18:00 265.5 F 25.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.2  
1:00 272.5 F 29.6  
2:00 271.4 F 29.4  
3:00 273.1 F 30.2  
4:00 271.7 F 29.7  
  
Min 271.0 F 29.0  
Max 273.1 F 30.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 198.6 F 0.0  
5:00 227.0 F 0.0  
10:00 236.7 F 0.0  
15:00 224.8 F 0.0  
20:00 222.1 F 0.0  
25:00 223.8 F 0.0  
30:00 223.7 F 0.0  
Drying Complete

FILLING: 1:36  
HEATING: 19:39  
STERILIZING: 4:00  
VENTING: 3:17  
DRYING: 30:00  
TOTAL CYCLE: 00:58:32  
CYCLE COMPLETE

03/22/2023 02:36 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_