

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 001386  
12:45 16/12/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:22  
PRESSURIZING 2:37  
STERILIZING 3:46  
136.5 C 220kPa 3:46  
136.0 C 220kPa 4:16  
135.8 C 220kPa 4:46  
136.2 C 223kPa 5:16  
136.4 C 223kPa 5:46  
136.2 C 224kPa 6:16  
136.4 C 225kPa 6:46  
136.3 C 224kPa 7:16  
136.3 C 224kPa 7:16  
VENTING 7:16  
AIR DRYING 8:07  
CYCLE COMPLETE 17:30  
-----

Digital Signature #  
A5478D1076EF3F48  
-----