

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4391

Name: RITTER2  
Sterilizer ID: V2313228  
02/28/2025 01:15 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 166.9 F 0.0  
2:00 165.2 F 2.3  
4:00 171.9 F 6.5  
6:00 196.1 F 11.5  
8:00 219.1 F 6.1  
10:00 226.0 F 11.4  
12:00 231.9 F 8.7  
14:00 239.0 F 10.5  
16:00 245.5 F 13.8  
18:00 257.8 F 20.4  
20:00 268.3 F 27.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.6  
1:00 271.9 F 29.6  
2:00 272.3 F 29.9  
3:00 272.3 F 30.1  
4:00 271.3 F 29.9  
  
Min 271.2 F 29.5  
Max 272.6 F 30.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 202.6 F 0.0  
5:00 232.7 F 0.0  
10:00 228.8 F 0.0  
15:00 229.1 F 0.0  
20:00 229.5 F 0.0  
25:00 228.7 F 0.0  
30:00 230.5 F 0.0  
Drying Complete

FILLING: 1:21  
HEATING: 21:07  
STERILIZING: 4:00  
VENTING: 3:42  
DRYING: 30:00  
TOTAL CYCLE: 01:00:11  
CYCLE COMPLETE

02/28/2025 02:15 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_